

CE DECLARATION OF CONFORMITY

NAME OF APPARATUS: *Electronic Vacuum Gauge*

MODEL NUMBER: *VG60 (DV150), VG64, VG64B*

DESCRIPTION OF APPARATUS: *Electronic Vacuum Gauge*

Sealed Unit Parts Co., Inc. declares that the above model is in conformity to the following harmonized standards and directives:

Directive 2006/95/EC		
Normative standard EN 61326-1: 2006	SPECIFICATIONS TO WHICH CONFORMITY IS DECLARED	LEVEL
EN 61000-4-2: 2008 Electrostatic Discharge Test	2,4,6,abd 8 kV Air Discharge 2,4, and 6 kV Contact Discharge	A
EN 61000-4-3: 2008 Radiated Susceptibility Test	4.5 V/M Minimum (4.5 V/M test 900 MHz Cell Phone, 200 Hz, 100% AM	A
EN 55011/EN55022 Radiated Emission	CISPR 16-2-3	B

Manufacturer:

Sealed Unit Parts Co.
P.O. Box 21
Allenwood, NJ 08720
USA

Manufacturer's Signature:



Alexander Brodetsky Ph.D
Sr.Vice President, Engineering

Date: 7/11/2016